



भारतीय प्रौद्योगिकी संस्थान इंदौर
खण्डवा रोड, सिमरोल, इंदौर - 453 552, भारत
Indian Institute of Technology Indore
Khandwa Road, Simrol, Indore - 453 552, India

www.iiti.ac.in

IIT Indore

Ref. No.: IITI(MM)/BSBE(PRJ)/1/1A/133/H CJ/2025-2026

December 24, 2025

PREBID REPORT

The online meeting for Pre-bid discussion was held at **IIT-Indore** through online channel on **18/12/2025 at 03:00 PM onwards** for GeM Custom bidding for **Supply and Installation of Benchtop High Performance Lyophilizer**.

Please find below the queries received and their responses.

Sl. No.	Reference of the Clause/ Page No. of the Tender Document	Query raised	Query Raised by	Response from IITI
1.	Page No. 09 Sr. No.: 10 Vacuum pump: Dry Scroll pump (Vacuubrand / Edward make) with suction rate 150 lit/min or better and Ultimate vacuum generated: Minimum 6.7 x104 mbar or better.	As per point no - 10 in specification file, it's said - dry scroll pump (vaccubrand/edward make) you have mentioned, but we can provide the equilant make vacuum pump which is better in pump specification, so we need confirmation why you need vaccubrand/edward make pump and if we have other brand pump, then we can quote or not.	Bidder through GeM Portal	Accepted Can be read as: Vacuum pump: Dry Scroll pump (Vacuubrand / Edward make - preferred) with suction rate 150 lit/min or better and Ultimate vacuum generated: Minimum 6.7 x104 mbar or better. Any make and model Dry Scroll pump can be quoted; however, bidder must submit user recommendation for organic samples / alkaloids etc at -105°C or better and installation details/order copy from Institute of National importance (IIT / IISER / CSIR / NIT etc.)

All prospective/willing bidders are requested to take note of this report as part of the Tender document. All other terms and conditions of the tender remain unchanged.

[Handwritten Signature]
24/12/25

Assistant Registrar
MM Section, IIT Indore

सहायक कुलसचिव
(सामग्री प्रबंधन विभाग)
Assistant Registrar
(Materials Management)